

## ANALYSIS CERTIFICATE

Tris NaCl BSA BUFFER - #AR005A / AR005K

Lot : F1700207

QC release : 2017-03-08

Expiration date : 2019-08-11

Reference AR005A: 1 x 50 ml

Reference AR005K: 3 x 50 ml

Components	Volume	Exp. (months)	Lot #	Exp. date
Tris NaCl BSA buffer	50 ml	30	F1700207	2019-08-11

503

## ANALYSIS CERTIFICATE

Tris NaCl BSA BUFFER - #AR005A / AR005K

Lot : F1700207

QC release: 2017-03-08

Expiration date : 2019-08-11

Analytical data	Specifications
<p><b>1. <u>Batch reproducibility</u></b></p> <p style="text-align: center;">N : 5 Volume &gt;50 mL</p>	≥ 50 mL
<p><b>2. <u>Aspect</u></b></p> <p style="text-align: center;">Clear, no sign of contamination</p>	Clear, no sign of contamination
<p><b>3. <u>VII-X clotting assay</u></b></p> <p style="text-align: center;">Instrument: STAR CT ( 100 %): 35 sec ΔCT (reference) 2,9 %</p>	CT (100%): 30 ± 5 sec ≤ 5%
<p><b>4. <u>Heparin anti-IIa chromogenic assay (EP):</u></b></p> <p style="text-align: center;">Instrument: Water Bath A405 (0 IU/ml) 1,299 ΔA405 (reference) 3,5 %</p>	≤ 5%

<b>Comments :</b>	<input checked="" type="checkbox"/> <b>PASSED IN COMPLIANCE</b>
-------------------	---

Date : 2017-03-08

QC Manager : S.LECOURT

