

## ANALYSIS CERTIFICATE

**FV deficient plasma - # DP020A/DP020K**

**Lot : F1700454**

**QC release: 2017-05-11**

**Expiration date : 2019-10-11**

**DP020A : 1 mL**

**DP020K : 6 x 1 mL**

<b>Components</b>	<b>Qty</b>	<b>Exp. (months)</b>	<b>LOT</b>	<b>Exp. date</b>
FV deficient plasma	Vial of 1 ml	30	F171100454	2019-10-11



## ANALYSIS CERTIFICATE

**FV deficient plasma - # DP020A/DP020K**

**Lot : F1700454**

**QC release: 2017-05-11**

**Expiration date : 2019-10-11**

Analytical data	Specifications
<b>1. <u>Within lot reproducibility</u></b>	
N= 10                      PT 57,4                      sec                      CV: 0,4 %	> 35 Sec (CV ≤ 3%)
N= 5                              aPTT >250                      sec	> 100 Sec

<b>2. <u>Concentration [C]</u></b>	
FV (ELISA)                      <1 %	< 1%

<b>3. <u>Aspect</u></b>	
a) :                      Slightly opalescent to clear	a) Slightly opalescent to clear
b) :                      No coagulum	b) No coagulum
c) :                      Stable solution	c) Stable solution

<b>4. <u>FV Clotting assay (MC10)</u></b>											
<table border="1" style="margin: auto;"> <tr> <td>V%</td> <td>100</td> <td>50</td> <td>25</td> <td>0</td> </tr> <tr> <td>CT (Sec)</td> <td>22,5</td> <td>27,4</td> <td>32,1</td> <td>65,4</td> </tr> </table>	V%	100	50	25	0	CT (Sec)	22,5	27,4	32,1	65,4	CT (0%) > 40 sec CT (100%) < 30sec
V%	100	50	25	0							
CT (Sec)	22,5	27,4	32,1	65,4							

<b>5. <u>Stability of reconstituted reagent</u></b>													
<table border="1" style="margin: auto;"> <tr> <td></td> <td style="text-align: center;">Fresh</td> <td style="text-align: center;">24h 2-8°C</td> <td style="text-align: center;">8h RT</td> </tr> <tr> <td>PT (sec)</td> <td style="text-align: center;">56,2</td> <td style="text-align: center;">56,1</td> <td style="text-align: center;">55,5</td> </tr> <tr> <td>CT (C%)(sec)</td> <td style="text-align: center;">23,5</td> <td style="text-align: center;">23,4</td> <td style="text-align: center;">23,7</td> </tr> </table>		Fresh	24h 2-8°C	8h RT	PT (sec)	56,2	56,1	55,5	CT (C%)(sec)	23,5	23,4	23,7	Δ ≤ 3 sec Δ ≤ 2 sec
	Fresh	24h 2-8°C	8h RT										
PT (sec)	56,2	56,1	55,5										
CT (C%)(sec)	23,5	23,4	23,7										

Comments :	<b>☑PASSED IN COMPLIANCE</b>
------------	----------------------------------

**Date :** 2017-05-11

**QC Manager :** S.LECOURT

