

Art no: 12-0105

CERTIFICATE OF ANALYSIS

Lot no: 224904



Product name:	USP-EP Anti-FXa & Anti-FIIa Assay for LMWH Buffer with BSA, pH 7.4
Production date:	2016-11
Date of release:	2016-12-14
Stability:	2019-11
Form:	White powder
Composition:	Contents of 1 pouch dissolved in deionized water and made up to 1000 ml yields: 0.050 M Tris buffer, pH 7.4 at 25°C 0.150 M NaCl 1.0 % (w/v) BSA (Bovine serum albumin)

Analysis	Specification	Result
Weight	26.35 g \pm 1 %	26.29 g (n=2), Pass
pH	7.40 \pm 0.10 at 25°C	7.36 (n=2), Pass
Optical clarity	Clear solution	Pass

The above material has met all quality specifications and has been reviewed by a quality representative.


Quality Assurance, Erik Silfverplatz

2016-12-14
Date

