



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN Factor XIII
Product Code : 227005
Lot No. : F1800294
Expiration date : 2019-11-03

Components	Quantity	Lot	Expiration Date
R1 : Thrombin	3 vials	F181400294	2019-11-03
R2 : Detection Reagent	3 vials	F181400294	2019-11-04

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	OD/min(100%) > 0.016 Delta OD/min (100-50%) ≥ 0.006 R ² ≥ 0.98 Measured values for controls in acceptance ranges	0,0215 0,0097 R ² = 1,00 controls are in compliance
Detection Threshold	≤ 5 %	≤ 1 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2018-03-12

Quality Control Manager : **M.PEYRAFITTE**