

ANALYSIS CERTIFICATE

Lyophilized platelets - #AG006A

Lot : F1700316
F1700675

QC Release: 2017-06-22

Expiration date : 2019-11-30

Components	Quantity	Lot #	Expiration date
Lyophilized Platelets	1 vial	1087-44	2019-11-30

825

ANALYSIS CERTIFICATE

Lyophilized platelets - #AG006A

**Lot : F1700316
 F1700675**

QC Release: 2017-06-22

Expiration date : 2019-11-30

Analytical data	Specifications
a. Lyophilized platelets	
Solubility of platelets : ≤ 10 minutes	≤ 10 minutes
Platelet Count : 280.10 ⁹ /L	N/A
Reference curve linearity : 0,97r ²	≥ 0.95r²
100% reference point : 59,8%/min	≥ 50 %/min.

Comments :	<input checked="" type="checkbox"/> Passed in Compliance
-------------------	---

Date : 2017-06-22

Redacted by : I. CORNUEJOLS QC Manager : S. LECOURT

