

# Fibrinogen Fragment E Fragment E du Fibrinogène

Vial of 200 µg/flacon de 200 µg

Ref. PP018A

STORE AT  
2-8 °C

Lot : 151119A

CONSERVER A  
2-8 °C

Expiration : 2019-11-30

## ANALYSIS CERTIFICATE

ANALYTICAL DATA	SPECIFICATIONS
1. <u>Protein Content</u> (Lowry method; A280 nm)  245 µg/vial	> 180 µg/vial
2. <u>SDS-PAGE</u> (5 % acrylamide)  1 single band of about 50000 daltons	1 major band of about 50,000 daltons

## CERTIFICAT D'ANALYSE

RESULTATS D'ANALYSE	SPECIFICATIONS
1. <u>Teneur protéique</u> (Méthode Lowry; D0280 nm)  245 µg/flacon	> 180 µg/flacon
2. <u>SDS-PAGE</u> (5 % acrylamide)  1 bande d'environ 50000 daltons	1 bande principale ≈ 50 000 daltons

DATE : 01/12/2015

CONCLUSIONS :  Passed/Conforme

Refused/Non-conforme

Quality Control Manager  
Responsable contrôle qualité      S. LECOURT

