



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN Plasminogen LRT
Product Code : 221511
Lot No. : F1701530-F1701531
Expiration date : 2019-12-02

Components	Quantity	Lot	Expiration Date
R1 : Streptokinase	3 vials	F171501530 F171201531	2019-12-02
R2 : Substrate	3 vials	F171501530 F171201531	2019-12-02

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	A405(100%) = 0.45 ± 0.20 Delta A405 (100-50%) > 0.15 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS
Detection Threshold	< 5 %	< 1 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2018-01-16

Quality Control Manager : **S.LECOURT**