



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN Plasminogen LRT  
**Product Code :** 221511  
**Lot No. :** F1701530-F1701531  
**Expiration date :** 2019-12-02

Components	Quantity	Lot	Expiration Date
R1 : Streptokinase	3 vials	F171501530 F171201531	2019-12-02
R2 : Substrate	3 vials	F171501530 F171201531	2019-12-02

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	A405(100%) = $0.45 \pm 0.20$ Delta A405 (100-50%) > 0.15 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS
Detection Threshold	< 5 %	< 1 %

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2018-01-16

Quality Control Manager : **S.LECOURT**