

## ANALYSIS CERTIFICATE

**IMIDAZOLE BUFFER - #AR021A / AR021K / AR021L**

**Lot : F1700712  
F1700713**

**QC release : 2017-07-17**

**Expiration date : 2019-12-14**

**Reference AR021A: 1 x 25 ml**

**Reference AR021K: 3 x 25 ml**

**Reference AR021L: 12 x 25 ml**

Components	Volume	Exp. (months)	Lot #	Exp. date
Imidazole buffer	25 ml	30	F171200712 F171100713	2019-12-14

925

## ANALYSIS CERTIFICATE

IMIDAZOLE BUFFER - #AR021A / AR021K / AR021L

Lot : F1700712  
 F1700713

QC release : 2017-07-17

Expiration date : 2019-12-14

Analytical data	Specifications
<b>1. <u>Batch reproducibility</u></b>	
N : 5	N ≥ 5
Volume >25 ml	≥ 25 ml
<b>2. <u>Aspect</u></b>	
<input checked="" type="checkbox"/> Clear, no sign of contamination	Clear, no sign of contamination
<b>3. <u>Hemoclot PC clotting assay</u></b>	
Instrument: MC10	
CT (25 %): 40,5 Sec.	CT (25%): 50 ± 10 sec
ΔCT(100%) (control sample/sample to test) 2,8 %	≤ 5%
<b>4. <u>FVIII:C clotting assay</u></b>	
Instrument: STAR	
CT (100%): 57,9 Sec.	CT (100%) < 75 sec
ΔCT (0-100%): 68,0 Sec.	ΔCT(0-100%) ≥ 40 sec
ΔCT(100%) (control sample/sample to test) 2 %	≤ 5%

<b>Comments :</b>	<input checked="" type="checkbox"/> <b>PASSED IN COMPLIANCE</b>
-------------------	---

Date : 2017-07-17

QC Manager : S. LECOURT

