

Purified Human Fibrinogen

Fibrinogène humain purifié

Vial of 1 g/flacon de 1 g

Ref. PP001C

STORE AT
2-8 °C

CONSERVER A
2-8 °C

Lot: F1600012

Expiration: 2019-12-23

ANALYSIS CERTIFICATE

ANALYTICAL DATA	SPECIFICATIONS
1. Protein Content (Lowry method) 1 049 mg/vial	> 900 mg/vial
2. SDS-PAGE (5 % acrylamide) 1 single band of about 340 000 daltons	1 major band of about 340,000 daltons
3. Clotting time (5 NIH/mL Thrombin) 7,7 sec	< 15 sec
4. Coagulability 98 %	≥ 98%

CERTIFICAT D'ANALYSE

RESULTATS D'ANALYSE	SPECIFICATIONS
1. Teneur protéique (Méthode Lowry) 1 049 mg/flacon	> 900 mg/flacon
2. SDS-PAGE (5 % acrylamide) 1 bande d'environ 340 000 daltons	1 bande principale ≈ 340 000 daltons
3. Temps de coagulation (Thrombine 5 NIH/mL) 7,7 sec	< 15 sec
4. Coagulabilité 98 %	≥ 98%

DATE : 04/02/2016

CONCLUSIONS : Passed/Conforme

Refused/Non-conforme

Quality Control Manager S. LECOURT
Responsable contrôle qualité

