

ANALYSIS CERTIFICATE

LIAPHEN AT (# 120002)

**Lot : F1700800
F1700801**

QC Release: 2017-11-06

Expiration date : 2020-01-04

Components	Qty	Exp. (months)	Lot	Exp. date
R1 : Latex reagent	2 vials	30	F171300800 F171100801	2020-03-28
R2 : Reaction Buffer	2 vials	30	F171300800 F171100801	2020-01-04

ANALYSIS CERTIFICATE

LIAPHEN AT (# 120002)

**Lot : F1700800
 F1700801**

QC Release: 2017-11-06

Expiration date : 2020-01-04

Analytical data	Specifications
<p>1. <u>Latex reagent (R1)</u> Method : Water bath</p> <p>Reproducibility (100% AT): N = 10 Mean : 0,412 CV: 0,7 %</p>	<p>≤ 3 %</p>
<p>2. <u>Reaction Buffer (R2)</u></p> <p style="text-align: right;">Volume: > 11 ml</p> <p style="text-align: right;">pH 8,08</p> <p>Aspect: <input checked="" type="checkbox"/> Clear <input checked="" type="checkbox"/> Transparent</p>	<p>≥ 10 ml</p> <p>[7.80 - 8.20]</p> <p>Clear - Transparent</p>

9/6

