

PLASMA PAI-1 CONTROL
Référence SC011K

Utilisation *in vitro* exclusivement

Uniquement à usage de recherche

FRANÇAIS

Lot : F1601054 - Exp. : 2020-03-30

PAI-1 Ag :

Control I lot : F1601055

Taux de PAI-1:Ag : 44 ng/ml

Intervalle de confiance : 37 à 51 ng/ml

Control II lot : F1601056

Taux de PAI-1:Ag : 16 ng/ml

Intervalle de confiance : 13 à 19 ng/ml

PAI-1 Activity :

Control I lot : F1601055

Taux de PAI-1 Activité : 12.1 ng/ml

Intervalle de confiance : 10.3 à 13.9 ng/ml

Control II lot : F1601056

Taux de PAI-1 Activité : 3.4 ng/ml

Intervalle de confiance : 2.7 à 4.1 ng/ml

PLASMA PAI-1 CONTROL
Référence SC011K

Utilisation *in vitro* exclusivement

Uniquement à usage de recherche

FRANÇAIS

Lot : F1601054 - Exp. : 2020-03-30

PAI-1 Ag :

Control I lot : F1601055

Taux de PAI-1:Ag : 44 ng/ml

Intervalle de confiance : 37 à 51 ng/ml

Control II lot : F1601056

Taux de PAI-1:Ag : 16 ng/ml

Intervalle de confiance : 13 à 19 ng/ml

PAI-1 Activity :

Control I lot : F1601055

Taux de PAI-1 Activité : 12.1 ng/ml

Intervalle de confiance : 10.3 à 13.9 ng/ml

Control II lot : F1601056

Taux de PAI-1 Activité : 3.4 ng/ml

Intervalle de confiance : 2.7 à 4.1 ng/ml

PLASMA PAI-1 CONTROL
SC011K

For *in vitro* use only

For research use only

ENGLISH

Lot : F1601054 - Exp. : 2020-03-30

PAI-1 Ag :

Control I lot : F1601055

PAI-1:Ag concentration : 44 ng/ml

Acceptance range : 37 to 51 ng/ml

Control II lot : F1601056

PAI-1:Ag concentration : 16 ng/ml

Intervalle de confiance : 13 to 19 ng/ml

PAI-1 Activity :

Control I lot : F1601055

PAI-1 Activity concentration : 12.1 ng/ml

Intervalle de confiance : 10.3 to 13.9 ng/ml

Control II lot : F1601056

PAI-1 Activity concentration : 3.4 ng/ml

Acceptance range : 2.7 to 4.1 ng/ml

PLASMA PAI-1 CONTROL
SC011K

For *in vitro* use only

For research use only

ENGLISH

Lot : F1601054 - Exp. : 2020-03-30

PAI-1 Ag :

Control I lot : F1601055

PAI-1:Ag concentration : 44 ng/ml

Acceptance range : 37 to 51 ng/ml

Control II lot : F1601056

PAI-1:Ag concentration : 16 ng/ml

Intervalle de confiance : 13 to 19 ng/ml

PAI-1 Activity :

Control I lot : F1601055

PAI-1 Activity concentration : 12.1 ng/ml

Intervalle de confiance : 10.3 to 13.9 ng/ml

Control II lot : F1601056

PAI-1 Activity concentration : 3.4 ng/ml

Acceptance range : 2.7 to 4.1 ng/ml

Quality Control Approval

17 NOV. 2016

S. Lecourt