



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN AT (Anti Ila)
Product Code : 221122
Lot No. : F1800037 / F1800038
Expiration date : 2020-07-05

Components	Quantity	Lot	Expiration Date
R1 : Bovin Thrombin	2 vials	F181300037 F181100038	2020-07-12
R2 : Substrate	2 vials	F181300037 F181100038	2020-07-11
R3: Tris heparin Buffer	2 vials	F181300037 F181100038	2020-07-05

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	Delta A405 (0-100%) ≥ 0.80 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS
Detection Threshold	$\leq 10\%$	< 2 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2018-02-14

Quality Control Manager : **S.LECOURT**