



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** Factor VIII:C Deficient Plasma  
**Product Code :** DP040A/K  
**Lot No. :** F1800373  
**Expiration date :** 2020-09-12

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	F181300373	2020-09-12

Analytical Data	Specification	Result
Calibration curve (STAR)	CT (0-100%) $\geq$ 40 sec CT (100%) $<$ 75 sec	57,9 53,4
Concentration FVIII	$<$ 1%	$<$ 1 %
Concentration FV	$>$ 50 %	85%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2018-04-10

Quality Control Manager : C.DUNOIS