



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** HEMOCLOT LA-C  
**Product Code :** CK091K  
**Lot No. :** F1800878  
**Expiration date :** 2020-11-20

Components	Quantity	Lot	Expiration Date
R1: HEMOCLOT LA-C	6 vials	F181200878	2020-11-20

Analytical Data	Specification	Result
Normal Plasma	Mean CT < 45 sec NR < 1.20	Mean CT = 24.6 sec (CS) Mean CT = 27.1 sec (STAR) NR = 0.99
Pathological Plasma	NR ≥ 1.20	NR > 1.20

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2018-07-16

Quality Control Manager :

M. PEYRAFITTE