



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** Factor IX Deficient Plasma  
**Product Code :** DP050A/K  
**Lot No. :** F1800921-F1800922  
**Expiration date :** 2020-12-01

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	F181300921 F181100922	2020-12-01

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) $\geq$ 40 sec CT (100%) $<$ 80 sec	107.6 sec 55.8 sec
Concentration FIX	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	90%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2018-06-25

Quality Control Manager :

M. PEYRAFITTE

P/O