



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-  
Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** IMIDAZOLE BUFFER  
**Product Code :** AR021A/K/L  
**Lot No. :** F1801191-F1801193  
**Expiration date :** 2021-01-29

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	1 vial (AR021A) 3 vials (AR021K) 12 vials (AR021L)	F181401191-F181201193	2021-01-29

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample $\leq$ 5%	2.2%
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	2.3%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

2018-08-28

Quality Control Manager :

M.PEYRAFITTE  
P.O. 