



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-
Oise
France

CERTIFICATE OF ANALYSIS

Product Description : IMIDAZOLE BUFFER
Product Code : AR021A/K/L
Lot No. : F1801191-F1801193
Expiration date : 2021-01-29

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	1 vial (AR021A) 3 vials (AR021K) 12 vials (AR021L)	F181401191-F181201193	2021-01-29

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample \leq 5%	2.2%
FVIII:C clotting assay	Delta CT (100%) with control sample \leq 5%	2.3%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

2018-08-28

Quality Control Manager :

M.PEYRAFITTE
P.O. 