



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN RIVAROXABAN CALIBRATOR LOW

**Product Code :** 226001

**Lot No. :** F1801287 - F1801288

**Expiration date :** 2021-03-09

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	F181301287 / F181101288	2021-03-09
CAL II	4 vials	F181301287 / F181101288	2021-03-09
CAL III	4 vials	F181301287 / F181101288	2021-03-09

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 ng/mL 53 ng/mL 112 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2018-10-15

Quality Control Manager : **M.PEYRAFITTE**



**BIOPHEN™ Rivaroxaban Calibrator Low**

**REF** 226001

**LOT** F1801287  2021-03-09

	<b>UNIT</b>	<b>TARGET VALUE</b>	<b>WHO STD</b>
<b>CAL I</b> <b>LOT</b> F181301287	ng/mL	0	N/A
<b>CAL II</b> <b>LOT</b> F181301287		53	
<b>CAL III</b> <b>LOT</b> F181301287		112	

*P. de S. 10.18*  
*amw*