



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : F1801472-F1801474

Expiration date : 2021-04-30

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	F181201472 / F181401474	2021-04-30

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A620 (0µg/ml) ≤ 0.15 Delta A620 (0-20µg/ml) ≥ 0.40 Controls measured within acceptance ranges R <sup>2</sup> ≥ 0.98	0.112 0.583 Controls in range R <sup>2</sup> = 0.99
Limit of Quantification	≤ 1 µg/ml (in the test dilution (1:200))	< 1 µg/mL
Hook effect	> 90 µg/ml (in the test dilution (1:200))	> 90 µg/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2018-12-13

Quality Control Manager : S.LECOURT