



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN ARIXTRA CONTROL PLASMA  
**Product Code :** 224001  
**Lot No. :** F1801608  
**Expiration date :** 2021-05-25

Components	Quantity	Lot	Expiration Date
C1	6 vials	F181301608	2021-05-25
C2	6 vials	F181301608	2021-05-25

Analytical Data	Specification	Result
Concentration	C1 : 0.30 -0.50 µg/ml C2 : 1.05 -1.35 µg/ml	0.42 µg/mL 1.22 µg/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-01-14

Quality Control Manager :

**S.LECOURT**



A Company of Sysmex Group

155 rue d'Eragny  
95000 Neuville-sur-Oise  
France  
Tél. : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36  
Web site : www.hyphen-biomed.com



## BIOPHEN® ARIXTRA® CONTROL

**REF** 224001

**LOT**

F1801608



2021-05-25

	<b>UNIT</b>	<b>TARGET VALUE</b>	<b>ACCEPTANCE RANGE</b>	<b>WHO STD</b>
<b>C1</b>	µg/mL	0.42	0.32 – 0.52	N/A
<b>LOT F181301608</b>				
<b>C2</b>		1.22	1.06 – 1.38	
<b>LOT F181301608</b>				