



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL C1
Product Code : 224101
Lot No. : F1801638-F1801639
Expiration date : 2021-05-29

Components	Quantity	Lot	Expiration Date
C1	12 vials	F181401638-F181101639	2021-05-29

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml	0.24 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-01-10

Quality Control Manager :

S.LECOURT



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél. : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN UFH CONTROL C1

REF 224101

LOT F1801638  2021-05-29

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C1 LOT F181401638	IU/mL	0.24	0.14 – 0.34	6 th International standard UFH 07/328