



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** Factor VII Deficient Plasma  
**Product Code :** DP030A/K  
**Lot No. :** F1801644  
**Expiration date :** 2021-05-29

Components	Quantity	Lot	Expiration Date
Factor VII Deficient Plasma	1 vial (DP030A) 6 vials (DP030K)	F181301644	2021-05-29

Analytical Data	Specification	Result
Calibration curve	CT (0%) > 50 sec CT (100%) < 30 sec	58.7 sec 24.5 sec
Concentration FVII	< 1%	< 1%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-01-08

Quality Control Manager : **S.LECOURT**