

Protein S Deficient Plasma Plasma Déficiant en Protéine S

Vial of 1 mL/flacon de 1 mL (DP110A)
Set of 6 x 1 mL / Coffret de 6 x 1 mL (DP110K)

REF DP110A/DP110K

ENGLISH / FRANÇAIS

STORE AT
2-8 °C

LOT

F1801573

EXP

2021-05-17

CONSERVER A
2-8 °C

CERTIFICATE OF ANALYSIS/CERTIFICAT D'ANALYSE

| ANALYTICAL DATA/RESULTATS D'ANALYSE | SPECIFICATIONS |
|--|----------------|
| PS: <1 % | < 1% |
| aPTT Ratio/Ratio TCA: 1.2 (Reagent/Réactif: Cephén LR) | < 1.5 |
| FV: 90 % | > 50 % |

DATE :2018-12-10

CONCLUSIONS : Passed/Conforme

Refused/Non-conforme

Quality Control Manager
Responsable contrôle qualité

S.LECOURT

