

# Prekallikrein Pool Pool de Prékallikréine

Vial of 2 mL/flacon de 2 mL or Set of 6x2mL/Coffret de 6x2mL

Ref. PP501B / PP501K

STORE AT  
2-8 °C

CONSERVER A  
2-8 °C

Lot: F1700635

Expiration: 2021-05-11

## ANALYSIS CERTIFICATE

## CERTIFICAT D'ANALYSE

ANALYTICAL DATA	SPECIFICATIONS
1. <b>Specificity:</b> Chromogenic assay  $\Delta$ A405 / 10 min: 0.001	< 0.010
2. <b>Activity:</b> Chromogenic assay following activation with PK activator at 32 IU/mL  $\Delta$ A405 / 10 min: 0.48	0.40 to 0.60
3. <b>Dose response curve with PK activation</b>  $R^2 = 0.99$	$R^2 > 0.98$

RESULTATS D'ANALYSE	SPECIFICATIONS
1. <b>Spécificité:</b> Dosage chromogénique  $\Delta$ A405 / 10 min: 0.001	< 0,010
2. <b>Activité:</b> Dosage chromogénique après activation avec l'activateur de la PK à 32 IU/mL  $\Delta$ A405 / 10 min: 0.48	0,40 à 0,60
3. <b>Courbe dose-réponse avec activation de la PK</b>  $R^2 = 0.99$	$R^2 > 0,98$

DATE : 2017-06-30

CONCLUSIONS :  Passed/Conforme  Refused/Non-conforme

Quality Control Manager  
Responsable contrôle qualité

S. LECOURT

