



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL LOW  
**Product Code :** 223701  
**Lot No. :** F1900169  
**Expiration date :** 2021-06-01

Components	Quantity	Lot	Expiration Date
CI	6 vials	F181801646	2021-06-01
CII	6 vials	F181501645	2021-06-01

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml CII : 0.35 -0.65 IU/ml	0.26 IU/mL 0.51 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-01-30

Quality Control Manager :

**S.LECOURT**




A Company of Sysmex Group

155 rue d'Eragny  
95000 Neuville-sur-Oise  
France  
Tél. : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36  
Web site : www.hyphen-biomed.com



### BIOPHEN LMWH CONTROL LOW

**REF** 223701

**LOT** F1900169  2021-06-01

	<b>UNIT</b>	<b>TARGET VALUE</b>	<b>ACCEPTANCE RANGE</b>	<b>WHO STD</b>
<b>CI</b> <b>LOT</b> F181801646	IU/mL	0.26	0.18 – 0.34	3 <sup>rd</sup> International standard LMWH 11/176
<b>CII</b> <b>LOT</b> F181501645		0.51	0.39 – 0.63	