



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW
Product Code : 223701
Lot No. : F1901133
Expiration date : 2021-06-01

Components	Quantity	Lot	Expiration Date
CI	6 vials	F181801646	2021-06-01
CII	6 vials	F191600916	2022-01-18

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml CII : 0.35 -0.65 IU/ml	0.26 IU/mL 0.51 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-09-13

Quality Control Manager :

S. LECOURT



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN LMWH CONTROL LOW

REF 223701

LOT F1901133  2021-06-01

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C I LOT F181801646	IU/mL	0.26	0.18 - 0.34	3 rd WHO international standard LMWH 11/176
C II LOT F191600916		0.51	0.39 - 0.63	