



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor VIII:C Deficient Plasma
Product Code : DP040A/K
Lot No. : F1801684
Expiration date : 2021-06-05

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	F181501684	2021-06-05

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) $<$ 75 sec	62.3 sec 54.4 sec
Concentration FVIII	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	64%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-01-08

Quality Control Manager : **S.LECOURT**