



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN AT

Item Code : 120002

Lot No. : F1900012

Expiration date : 2021-07-02

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	2 vials	F191500012	2021-08-06
R2 : Reaction Buffer	2 vials	F191500012	2021-07-02

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A620 (0%) \leq 0.15 A620 (0-150%) \geq 0.40 R ² \geq 0.98 Controls measured within acceptance ranges	0.136 0.884 R ² = 1.000 Controls measured within acceptance ranges
Detection threshold	\leq 10%	< 1%
Hook effect	>200%	> 300%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-03-05

Quality Control Manager : **S.LECOURT**