



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : **COLLAGEN**
Product Code : **AG005K**
Lot No. : **F1900121-F1900122**
Expiration date : **2021-07-23**

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	F1900121-F1900122	2021-07-27
COLLAGEN DILUENT	3 vials	F1900121-F1900122	2021-07-23

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 10 µg/ mL	≥ 70%	83
Reactivity (% AgregMax) at 2 µg/mL	≥ 70%	89

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-03-08

Quality Control Manager : **S.LECOURT**

