

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description :	PLASMA HIRUDIN STANDARD HIGH
Product Code :	SC020L
Lot No. :	F1900119-F1900120
Expiration date :	2021-07-19


Components	Quantity	Lot	Expiration Date
STD 0 : Normal Plasma about 0 µg/mL Hir.	3 vials	F191200119/F191200120	2021-07-19
STD 5 : Normal Plasma about 5 µg/mL Hir.	3 vials	F191200119/F191200120	2021-07-19

Analytical Data	Specification	Result
STD 0	< 0.1 µg/mL	0 µg/mL
STD 5	5 ± 0.4 µg/mL	4.80 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-02-21

Quality Control Manager : **S.LECOURT**



PLASMA HIRUDIN STANDARD HIGH

REF SC020L-RUO

LOT F1900119  2021-07-19

	UNIT	TARGET VALUE	WHO STD
<p>STD 0</p> <p>LOT F191200119</p>	<p>µg/mL</p>	<p>0</p>	<p>NA</p>
<p>STD 5</p> <p>LOT F191200119</p>		<p>4.80</p>	