



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR
Product Code : 222701
Lot No. : F1900316
Expiration date : 2021-08-30


Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F191300316	2021-08-30
CAL 2	4 vials	F191300316	2021-08-30
CAL 3	4 vials	F191300316	2021-08-30

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 255 ng/mL 527 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	$R^2 > 0.98$ Measured values controls in acceptance ranges

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-04-05

Quality Control Manager : **S.LECOURT**





A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France

Tel : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN™ Rivaroxaban Plasma Calibrator

REF 222701

LOT F1900316  2021-08-30

LOT F1900316

	UNIT	TARGET VALUE	WHO STD	
CAL 1	ng/mL	0	N/A	
LOT F191300316		255		
CAL 2		527		
LOT F191300316				