



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL LOW
Product Code : 225101
Lot No. : F1900193
Expiration date : 2021-08-07

Components	Quantity	Lot	Expiration Date
C1	6 vials	F191500193	2021-08-07
C2	6 vials	F191500193	2021-08-07

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	25 ng/mL 80 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-03-11

Quality Control Manager :

S.LECOURT



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tel : +33.1.34.40.65.10 ... Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN™ Rivaroxaban Control Low

REF 225101

LOT F1900193  2021-08-07

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
[CI] [LOT] F191500193	ng/mL	25	[15 – 35]	N/A
[CII] [LOT] F191500193		80	[64 – 96]	