



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description :	Factor IX Deficient Plasma
Product Code :	DP050A/K
Lot No. :	F1900236
Expiration date :	2021-08-14

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	F191700236	2021-08-14

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) $<$ 80 sec	93.8 sec 56.6 sec
Concentration FIX	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	102%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-03-07

Quality Control Manager : **S.LECOURT**

