

HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS


**Product Description :** BIOPHEN RIVAROXABAN CALIBRATOR LOW  
**Product Code :** 226001  
**Lot No. :** F1900444  
**Expiration date :** 2021-09-25

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	F191400444	2021-09-25
CAL II	4 vials	F191400444	2021-09-25
CAL III	4 vials	F191400444	2021-09-25

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 52 102
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.00 Controls are in acceptance range

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-05-02      Quality Control Manager : **S.LECOURT**





A Company of Sysmex Group

155 rue d'Eragny  
95000 Neuville-sur-Oise  
France  
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36  
Web site : www.hyphen-biomed.com



### BIOPHEN™ Rivaroxaban Calibrator Low

**REF** 226001

**LOT** F1900444  2021-09-25

	<b>UNIT</b>	<b>TARGET VALUE</b>	<b>WHO STD</b>
<b>CAL I</b>  <b>LOT</b> F191400444	ng/mL	0	N/A
<b>CAL II</b>  <b>LOT</b> F191400444		52	
<b>CAL III</b>  <b>LOT</b> F191400444		102	