



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** HEMOCLOT LA-S  
**Product Code :** CK090K  
**Lot No. :** F1900347  
**Expiration date :** 2021-09-03

Components	Quantity	Lot	Expiration Date
R1: HEMOCLOT LA-S	6 vials	F191300347	2021-09-03

Analytical Data	Specification	Result
Normal Plasma	Mean CT < 45 sec NR < 1.20	30.0 sec 1.04
Pathological Plasma	NR ≥ 1.20	1.87

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-03-29

Quality Control Manager : S.LECOURT