



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** CEPHEN 1  
**Product Code :** CK511K  
**Lot No. :** F1901098  
**Expiration date :** 2021-09-05

Components	Quantity	Lot	Expiration Date
CEPHEN 1	6 vials	F193M01098	2021-09-05

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	0.92

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-10-16

Quality Control Manager :

S. LECOURT