



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ARGATROBAN CONTROL PLASMA  
**Product Code :** SC035K  
**Lot No. :** F1900442  
**Expiration date :** 2021-09-28

Components	Quantity	Lot	Expiration Date
C1	6 vials	F191300442	2021-09-28
C2	6 vials	F191300442	2021-09-28

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.70 µg/mL
C2	1.10 - 1.40 µg/mL	1.29 µg/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-04-29

Quality Control Manager : S.LECOURT



A Company of Sysmex Group  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France  
Tel : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36  
Web site : www.hyphen-biomed.com



## ARGATROBAN® CONTROL PLASMA

REF SC035K

LOT F1900442 2021-09-28

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C1 LOT F191300442	µg/mL	0.70	0.55 – 0.85	N/A
C2 LOT F191300442		1.29	1.09 – 1.49	