



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL C3  
**Product Code :** 223801  
**Lot No. :** F1900542  
**Expiration date :** 2021-10-19

Components	Quantity	Lot	Expiration Date
C3	12 vials	F191500542	2021-10-19

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml	0.81 IU/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-05-21


Quality Control Manager :

**S.LECOURT**




A Company of Sysmex Group

155 rue d'Eragny  
95000 Neuville-sur-Oise  
France  
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36  
Web site : [www.hyphen-biomed.com](http://www.hyphen-biomed.com)



### BIOPHEN LMWH CONTROL C3

**REF** 223801

**LOT** F1900542  2021-10-19

	<b>UNIT</b>	<b>TARGET VALUE</b>	<b>ACCEPTANCE RANGE</b>	<b>WHO STD</b>
<b>C3</b> <b>LOT</b> F191500542	IU/mL	0.81	0.69 – 0.93	11/176