



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL C1
Product Code : 224101
Lot No. : F1900549
Expiration date : 2021-10-23

Components	Quantity	Lot	Expiration Date
C1	12 vials	F191500549	2021-10-23

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml	0.25 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-05-27

Quality Control Manager :

S.LECOURT



A Company of Sysmex Group

155 rue d'Eragny

95000 Neuville-sur-Oise

France

Tél. : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36

Web site : www.hyphen-biomed.com

BIOPHEN UFH CONTROL C1

REF 224101

LOT

F1900549



2021-10-23

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C1 LOT F191500549	IU/mL	0.25	0.15 – 0.35	07/328