

Art no: 12-0018

CERTIFICATE OF ANALYSIS

Lot no: 224702



Product name:	USP-EP Anti-FXa & Anti-FIIa Assay for UFH Buffer with PEG, pH 8.4
Production date:	2016-11
Date of release:	2016-12-14
Stability:	2021-11
Form:	White powder
Composition:	Contents of 1 pouch dissolved in deionized water and made up to 1000 ml yields: 0.050 M Tris buffer, pH 8.4 at 25°C 0.175 M NaCl 0.0075 M EDTA 0.10% (w/v) PEG 6000

Analysis	Specification	Result
Weight	20.4 g \pm 1 %	20.3 (n=2), Pass
pH	8.40 \pm 0.10 at 25°C	8.41 (n=3), Pass
Optical clarity	Visually clear solution	Pass

The above material has met all quality specifications and has been reviewed by a quality representative.


Quality Assurance, Erik Silfverplatz

2016-12-14
Date