



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR

Product Code : 226201

Lot No. : F1900693

Expiration date : 2021-11-30

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F191500693	2021-11-30
CAL 2	4 vials	F191500693	2021-11-30
CAL 3	4 vials	F191500693	2021-11-30


Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 500 -700 ng/ml	0 303 601
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.99 controls in acceptance range

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-07-08

Quality Control Manager :

S.LECOURT





A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN™ Apixaban Calibrator

REF 226201

LOT F1900693

 2021-11-30

CAL 1	CAL 2	CAL 3	UNIT	TARGET VALUE	WHO STD
LOT F191500693	LOT F191500693	LOT F191500693	ng/mL	0	NA
				303	
				601	