



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** Factor VIII:C Deficient Plasma  
**Product Code :** DP040A/K  
**Lot No. :** F1900785  
**Expiration date :** 2021-12-20

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	F191600785	2021-12-20

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) $\geq$ 40 sec CT (100%) $<$ 75 sec	72.0 sec 56.1 sec
Concentration FVIII	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	69%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2019-07-12

Quality Control Manager : **S.LECOURT**