



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW CII
Product Code : 224401
Lot No. : F1900916
Expiration date : 2022-01-18

Components	Quantity	Lot	Expiration Date
CII	12 vials	F191600916	2022-01-18

Analytical Data	Specification	Result
Concentration	CII : 0.35 -0.65 IU/ml	0.51

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-08-30


Quality Control Manager :

S. LECOURT



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tel : +33.1.34.40.65.10 – Fax : +33.1.34.46.72.36
Web site : www.hyphen-biomed.com



BIOPHEN LMWH CONTROL LOW CII

REF 224401

LOT F1900916  2022-01-18

CII LOT F191600916	UNIT IU/mL	TARGET VALUE 0.51	ACCEPTANCE RANGE [0.39 – 0.63]	WHO STD 3 rd WHO international standard LMWH 11/176
-------------------------------------	--------------------------	---------------------------------	--	--

Approved by 30/08/2023
P.O. GUYOT