



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code : 224501
Lot No. : F1900968
Expiration date : 2022-01-22

Components	Quantity	Lot	Expiration Date
C1	6 vials	F191600968	2022-01-22
C2	6 vials	F191600968	2022-01-22

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	99 ng/ml 292 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-09-11

Quality Control Manager :

S. LECOURT

BIOPHEN™ Rivaroxaban Control Plasma
REF 224501

LOT F1900968

 2022-01-22

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C1 LOT F191600968	ng/mL	99 (ng)	69 - 129	N/A
C2 LOT F191600968	ng/mL	292	245 - 339	N/A