



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN
Product Code : AG004K
Lot No. : F1900908
Expiration date : 2022-01-10

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	F1900908	2022-01-10

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.5 mg/mL	≥ 80%	95%
Reactivity (% AgregMax) at 1 mg/mL	≥ 70%	92%
Reactivity (% AgregMax) at 0.5 mg/mL	< 10%	2%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-08-30

Quality Control Manager : **S. LECOURT**

