

CERTIFICATE OF ANALYSIS

Product Description :	PLASMA ARGATROBAN CALIBRATOR
Product Code :	SC030K
Lot No. :	F1900867
Expiration date :	2022-01-03


Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	F191300867	2022-01-03
CAL 2	4 vials	F191300867	2022-01-03
CAL 3	4 vials	F191300867	2022-01-03
CAL 4	4 vials	F191300867	2022-01-03
CAL 5	4 vials	F191300867	2022-01-03

Analytical Data	Specification	Result
CAL 1	≤ 0.05 µg/mL	0
CAL 2	0.35 - 0.65 µg/mL	0.53
CAL 3	0.85 - 1.15 µg/mL	1.06
CAL 4	1.30 - 1.65 µg/mL	1.61
CAL 5	1.80 - 2.10 µg/mL	2.06

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-09-16

Quality Control Manager : **S. LECOURT**



ARGATROBAN® PLASMA CALIBRATOR

REF SC030K

LOT F-1900867 2022-01-03

UNIT	TARGET VALUE	WHO STD
CAL 1	0	N/A
LOT F191300867		
CAL 2	0.53	
LOT F191300867		
CAL 3	1.06	
LOT F191300867		
CAL 4	1.61	
LOT F191300867		
CAL 5	2.06	
LOT F191300867		