



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CONTROL LOW
Product Code : 225401
Lot No. : F1900993
Expiration date : 2022-02-08

Components	Quantity	Lot	Expiration Date
CI	6 vials	F191300993	2022-02-08
CII	6 vials	F191300993	2022-02-08

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -105 ng/ml	24 ng/ml 74 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-09-16

Quality Control Manager : S. LECOURT



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN™ Edoxaban Control Low

REF 225401

LOT F1900993  2022-02-08

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
CI LOT F191300993	ng/mL	24	14 - 34	N/A
CII LOT F191300993		74	59 - 89	