



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL C3
Product Code : 223801
Lot No. : F1901204
Expiration date : 2022-03-28

Components	Quantity	Lot	Expiration Date
C3	12 vials	F191701204	2022-03-28

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml	0.81 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-11-21

Quality Control Manager :

C.DUNOIS



A Company of Sysmex Group

155 rue d'Eragny
95000 Neuville-sur-Oise
France
Tél. : +33.1.34.40.65.10 – Fax : +33.1.34.48.72.36
Web site : www.hyphen-biomed.com



BIOPHEN LMWH CONTROL C3

REF 223801

LOT F1901204  2022-03-28

	UNIT	TARGET VALUE	ACCEPTANCE RANGE	WHO STD
C3 LOT F191701204	IU/mL	0.81	0.69 - 0.93	3rd International standard for LMWH 11/176