




A Company of Sysmex Group

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : F1903092

Expiration date : 2022-04-24

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	F191703092	2022-04-24

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A620 (0µg/ml) ≤ 0.15 Delta A620 (0-20µg/ml) ≥ 0.40 Controls measured within acceptance ranges R ² ≥ 0.98	0,15 0,56 controls in range 0,99
Limit of Quantification	≤ 1 µg/ml (in the test dilution (1:200))	0,25
Hook effect	> 90 µg/ml (in the test dilution (1:200))	140

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2019-12-02

Quality Control Manager : S.LECOURT

